

DECLARATION OF CONFORMITY

According to ISO/IEC 17050-1:2010

Manufacturer's Name: Crystal Engineering Corporation
An AMETEK Inc. company

Manufacturer's Address: 708 Fiero Lane, Suite 9
San Luis Obispo, CA 93401
USA

Declares under sole responsibility that the product as originally delivered

Product Name: Digital Pressure Gauge
Model Number: M1 Series

Complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

EMC Directive 2004/108/EC

Standard
CISPR 11:2003

EHSR
CISPR 11:2009/A1:2010 harmonized standard has been compared to the standard used for certification purposes and no changes in the "state of the art" apply to the equipment. (1)

EN 61326: 2006


Harmonized

And conforms with the following product standards:

Standard	Description	Class	Status	EHSR
CISPR 11:2003	Radiated Emissions	Class A	Pass	See (1)
EN 61326-1:2006 (EN 61000-4-2: 2009)	Electrostatic Discharge	Criteria A	Pass	Harmonized
EN 61326-1:2006 (EN 61000-4-3: 2006)	Radiated Immunity	Criteria A	Pass	Harmonized
EN 61326-1:2006 (EN 61000-4-8: 2001)	Magnetic Field Immunity	Criteria A	Pass	Harmonized


I hereby declare that the equipment named above has been designed to comply with the relevant sections of the above referenced specifications. The unit complies with all essential requirements of the Directives

USA Signatory:



Division Vice President & Crystal Business Manager
David K Porter, P.E.

European Signatory:



Division Vice President & Business Unit Manager
Joel Frie

Crystal Engineering Corporation, an AMETEK Inc. company
708 Fiero Lane, Suite 9, San Luis Obispo, CA 93401, USA
+1 805 595 5477

AMETEK Denmark A/S
Gydevang 32-34, 3450 Allerød, Denmark
+45 4816 8000

11 March 2014